



Iss.	1	Date	21.09 2023	Modification	P0182. New drawing.	Ckd.		Appd.	
Iss.		Date		Modification		Ckd.		Appd.	
<b>BEKA</b> <b>associates</b> <b>England</b> Hitchin company confidential, copyright reserved.									




THIRD ANGLE PROJECTION

DIMENSIONS IN mm

DO NOT SCALE




VIEW ON ARROW A  
(REAR PANEL DETAILS)




#### TERMINAL CONNECTIONS:

OPTION TERMINALS (WHEN OPTION FITTED):



#### INSTALLATION:

- 1 INSERT UNIT INTO PANEL CUTOUT FROM FRONT.
- 2 SCREW CLAMPS ARE FITTED OVER THE 4 STUDS & THE WING NUTS TIGHTENED FROM REAR OF PANEL. RECOMMENDED TORQUE 22cNm MIN.
- 3 FOR PANELS GREATER THAN 3mm THICK ADD ADDITIONAL DEPTH FOR CLAMPS. (ie ADD PANEL THICKNESS LESS 3mm)
- 4 FINALLY, FIT A SILICONE RUBBER CAP TO THE END OF EACH M3 THREADED ROD.



Title  
CUSTOMER DRAWING for  
BA307SE 4 DIGIT INDICATOR  
IN PANEL MOUNTING ENCLOSURE

Drawn KB	Checked	Scale NTS
Drawing No. CD307SE		